

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
99.845 ml	100 ml	0.155 ml	± 1.0 ml	±	1.4E-01 µl
499.51 ml	500 ml	0.488 ml	± 5.0 ml	±	1.4E-01 µl
1,001.31 ml	1000 ml	-1.311 ml	± 10.0 ml	±	1.5E-01 µl