

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
50.074 ml	50 ml	-0.074 ml	± 1.0 ml	±	1.4E-01 µl
249.96 ml	250 ml	0.038 ml	± 2.0 ml	±	1.4E-01 µl
499.85 ml	500 ml	0.147 ml	± 5.0 ml	±	1.4E-01 µl