

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
10.013 ml	10 ml	-0.013 ml	± 0.1 ml	±	3.1E-01 µl
149.90 ml	150 ml	0.095 ml	± 1.0 ml	±	4.6E+00 µl
249.99 ml	250 ml	0.008 ml	± 2.0 ml	±	7.6E+00 µl