

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
12 °C	12.314 °C	12.1 °C	-0.214 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.255 °C	27.0 °C	-0.255 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.146 °C	42.1 °C	-0.046 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.043 %HR	51.6 %HR	-0.443 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	72.186 %HR	71.8 %HR	-0.386 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.329 %HR	90.6 %HR	-1.729 %HR	± 3.0 %HR	± 1.9E+00 %HR