

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 12 °C             | 12.314 °C         | 12.1 °C   | -0.214 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 27 °C             | 27.255 °C         | 27.1 °C   | -0.155 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 42 °C             | 42.146 °C         | 42.0 °C   | -0.146 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 52 %HR            | 52.043 %HR        | 50.5 %HR  | -1.543 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 72 %HR            | 72.186 %HR        | 71.9 %HR  | -0.286 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| # 92 %HR          | 92.329 %HR        | 90.4 %HR  | -1.929 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |