

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.270 °C        | -24.7 °C  | 0.570 °C      | ± 0.7 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.251 °C         | -0.2 °C   | 0.051 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 4.733 °C          | 4.7 °C    | -0.033 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 9.736 °C          | 9.7 °C    | -0.036 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 24.888 °C         | 24.9 °C   | 0.012 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 29.902 °C         | 29.6 °C   | -0.302 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 39.909 °C         | 39.6 °C   | -0.309 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |

