

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.270 °C	-25.0 °C	0.270 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.251 °C	-0.2 °C	0.051 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.733 °C	4.9 °C	0.167 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.736 °C	9.9 °C	0.164 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.888 °C	25.3 °C	0.412 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.902 °C	30.3 °C	0.398 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.909 °C	40.4 °C	0.491 °C	± 0.5 °C	± 4.0E-01 °C

