

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.270 °C	-25.4 °C	-0.130 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.251 °C	-0.6 °C	-0.349 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.733 °C	4.4 °C	-0.333 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.736 °C	9.5 °C	-0.236 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.888 °C	24.6 °C	-0.288 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.902 °C	29.7 °C	-0.202 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.909 °C	39.7 °C	-0.209 °C	± 0.5 °C	± 4.0E-01 °C

