

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.181 °C	-0.1 °C	0.081 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	19.995 °C	20.0 °C	0.005 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.902 °C	30.0 °C	0.098 °C	± 0.5 °C	± 4.0E-01 °C
50 °C	50.057 °C	50.0 °C	-0.057 °C	± 0.5 °C	± 1.1E+00 °C