

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.621 °C	-25.5 °C	0.121 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.271 °C	0.0 °C	0.271 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.773 °C	5.2 °C	0.427 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.827 °C	10.3 °C	0.473 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.928 °C	25.2 °C	0.272 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.882 °C	30.2 °C	0.318 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.960 °C	40.4 °C	0.440 °C	± 0.5 °C	± 4.0E-01 °C

