

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.621 °C	-25.5 °C	0.121 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.271 °C	-0.5 °C	-0.229 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.773 °C	4.7 °C	-0.073 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.827 °C	9.7 °C	-0.127 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.928 °C	24.8 °C	-0.128 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.882 °C	29.9 °C	0.018 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.960 °C	39.7 °C	-0.260 °C	± 0.5 °C	± 4.0E-01 °C

