

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.621 °C        | -25.7 °C  | -0.079 °C     | ± 0.7 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.271 °C         | -0.6 °C   | -0.329 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 4.773 °C          | 4.6 °C    | -0.173 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 9.827 °C          | 9.7 °C    | -0.127 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 24.928 °C         | 24.6 °C   | -0.328 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 29.882 °C         | 29.9 °C   | 0.018 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 39.960 °C         | 39.9 °C   | -0.060 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |

