

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
19,897.7 lx	19,892.0 lx	5.7 lx	± 498 lx	±	9.8E+01 lx
1,035.5 lx	1,034.0 lx	1.5 lx	± 27 lx	±	1.1E+00 lx
340.7 lx	339.4 lx	1.3 lx	± 9 lx	±	1.8E-01 lx
30.2 lx	28.9 lx	1.3 lx	± 2 lx	±	2.9E-02 lx