

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	Cumplimiento
100.005 ml	100 ml	-0.005 ml	± 5.0 ml	±1.4E-01 ml	Cumple
249.963 ml	250 ml	0.037 ml	± 5.0 ml	±1.4E-01 ml	Cumple
499.907 ml	500 ml	0.093 ml	± 5.0 ml	±1.4E-01 ml	Cumple
750.061 ml	750 ml	-0.061 ml	± 5.0 ml	±1.4E-01 ml	Cumple
1,000.114 ml	1000 ml	-0.114 ml	± 5.0 ml	±1.5E-01 ml	Cumple
Según especificaciones del fabricante. EMP: ± 5 ml					