

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	Cumplimiento
100.015 ml	100 ml	-0.015 ml	± 5.0 ml	±1.4E-01 µl	Cumple
250.027 ml	250 ml	-0.027 ml	± 5.0 ml	±1.4E-01 µl	Cumple
499.918 ml	500 ml	0.082 ml	± 5.0 ml	±1.4E-01 µl	Cumple
750.073 ml	750 ml	-0.073 ml	± 5.0 ml	±1.4E-01 µl	Cumple
1,000.088 ml	1000 ml	-0.088 ml	± 5.0 ml	±1.5E-01 µl	Cumple
Según especificaciones del fabricante. EMP: ± 5 ml					