

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
99.988 ml	100 ml	0.012 ml	± 5.0 ml	±	1.4E-01 µl
250.037 ml	250 ml	-0.037 ml	± 5.0 ml	±	1.4E-01 µl
499.969 ml	500 ml	0.031 ml	± 5.0 ml	±	1.4E-01 µl
750.067 ml	750 ml	-0.067 ml	± 5.0 ml	±	1.4E-01 µl
1,000.052 ml	1000 ml	-0.052 ml	± 5.0 ml	±	1.5E-01 µl