

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|------------|
| 99.996 ml                | 100 ml           | 0.004 ml             | ± 5.0 ml                    | ±                                  | 1.4E-01 µl |
| 250.023 ml               | 250 ml           | -0.023 ml            | ± 5.0 ml                    | ±                                  | 1.4E-01 µl |
| 499.980 ml               | 500 ml           | 0.020 ml             | ± 5.0 ml                    | ±                                  | 1.4E-01 µl |
| 750.078 ml               | 750 ml           | -0.078 ml            | ± 5.0 ml                    | ±                                  | 1.4E-01 µl |
| 1,000.065 ml             | 1000 ml          | -0.065 ml            | ± 5.0 ml                    | ±                                  | 1.5E-01 µl |