

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	Cumplimiento
99.996 ml	100 ml	0.004 ml	± 5.0 ml	±1.4E-01 ml	Cumple
250.023 ml	250 ml	-0.023 ml	± 5.0 ml	±1.4E-01 ml	Cumple
499.980 ml	500 ml	0.020 ml	± 5.0 ml	±1.4E-01 ml	Cumple
750.078 ml	750 ml	-0.078 ml	± 5.0 ml	±1.4E-01 ml	Cumple
1,000.065 ml	1000 ml	-0.065 ml	± 5.0 ml	±1.5E-01 ml	Cumple
Según especificaciones del fabricante. EMP: ± 5 ml					