

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	Cumplimiento
99.978 ml	100 ml	0.022 ml	± 5.0 ml	±1.4E-01 µl	Cumple
250.026 ml	250 ml	-0.026 ml	± 5.0 ml	±1.4E-01 µl	Cumple
499.946 ml	500 ml	0.054 ml	± 5.0 ml	±1.4E-01 µl	Cumple
750.100 ml	750 ml	-0.100 ml	± 5.0 ml	±1.4E-01 µl	Cumple
999.991 ml	1000 ml	0.009 ml	± 5.0 ml	±1.5E-01 µl	Cumple

Según especificaciones del fabricante. EMP: ± 5 ml