

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
9.980 ml	10 ml	0.020 ml	± 0.5 ml	±	3.1E-01 µl
25.004 ml	25 ml	-0.004 ml	± 1.0 ml	±	7.8E-01 µl
50.044 ml	50 ml	-0.044 ml	± 2.5 ml	±	1.5E+00 µl