

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |             |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|-------------|
| 9.980 ml                 | 10 ml            | 0.020 ml             | ± 0.50 ml                   | ±                                  | ±3.1E-01 ml |
| 25.004 ml                | 25 ml            | -0.004 ml            | ± 0.50 ml                   | ±                                  | ±7.8E-01 ml |
| 50.044 ml                | 50 ml            | -0.044 ml            | ± 0.50 ml                   | ±                                  | ±1.5E+00 ml |

Según especificaciones del fabricante. EMP: ± 0.5 ml