

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	Cumplimiento
100.078 ml	100 ml	-0.078 ml	± 1.0 ml	±1.4E-01 µl	Cumple
149.908 ml	150 ml	0.092 ml	± 1.0 ml	±1.4E-01 µl	Cumple
200.023 ml	200 ml	-0.023 ml	± 1.0 ml	±1.4E-01 µl	Cumple

Según especificaciones del fabricante. EMP: ± 1 ml