

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	Cumplimiento
4.998 ml	5 ml	0.002 ml	± 0.20 ml	±1.7E-01 µl	Cumple
9.922 ml	10 ml	0.078 ml	± 0.20 ml	±3.1E-01 µl	Cumple
50.024 ml	50 ml	-0.024 ml	± 0.20 ml	±1.5E+00 µl	Cumple

Según especificaciones del fabricante. EMP: ± 0.2 ml