

Certificado No. (Certificate #):63276Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-11-14Marca (Brand):Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
0.05000 in	0.05010 in	0.00010 in	± 0.0001 in	±	3.1E-05 in
0.10050 in	0.10055 in	0.00005 in	± 0.0001 in	±	3.2E-05 in
0.10100 in	0.10105 in	0.00005 in	± 0.0001 in	±	3.3E-05 in
0.10500 in	0.10505 in	0.00005 in	± 0.0001 in	±	3.3E-05 in
0.11000 in	0.11010 in	0.00010 in	± 0.0001 in	±	3.2E-05 in
0.15000 in	0.15010 in	0.00010 in	± 0.0001 in	±	3.3E-05 in
0.20000 in	0.20010 in	0.00010 in	± 0.0001 in	±	3.3E-05 in
0.30000 in	0.30010 in	0.00010 in	± 0.0001 in	±	3.4E-05 in
0.50000 in	0.50010 in	0.00010 in	± 0.0001 in	±	3.4E-05 in
0.80000 in	0.80010 in	0.00010 in	± 0.0001 in	±	3.4E-05 in
1.00000 in	1.00010 in	0.00010 in	± 0.0001 in	±	3.3E-05 in