

Certificado No. (Certificate #):		63277		Equipo (Instrument):		Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-11-14		Marca (Brand):		Mitutoyo	
Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)		
Exterior / (Exterior)	1 in	0.5000 in	0.0000 in	± 0.0011811 in	± 9.2E-02 in		
Exterior / (Exterior)	1 in	1.0000 in	0.0000 in	± 0.0011811 in	± 3.3E-01 in		
Exterior / (Exterior)	3 in	3.0000 in	0.0000 in	± 0.0011811 in	± 4.2E-01 in		
Exterior / (Exterior)	4 in	4.0000 in	0.0000 in	± 0.0011811 in	± 5.0E-01 in		
Exterior / (Exterior)	8 in	8.0005 in	0.0005 in	± 0.0015748 in	± 6.1E-01 in		
Interior / (Interior)	2 in	2.0000 in	0.0000 in	± 0.0011811 in	± 2.3E-01 in		
Interior / (Interior)	4 in	4.0000 in	0.0000 in	± 0.0011811 in	± 4.2E-01 in		
Profundidad / (Depth)	2 in	2.0000 in	0.0000 in	± 0.0011811 in	± 2.4E-01 in		
Profundidad / (Depth)	4 in	4.0000 in	0.0000 in	± 0.0011811 in	± 3.5E-01 in		