

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.158 °C	10.1 °C	-0.058 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	19.955 °C	19.9 °C	-0.055 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.012 °C	29.9 °C	-0.112 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	48.055 %HR	49.8 %HR	1.745 %HR	± 3.0 %HR	± 1.9E+00 %HR