

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.621 °C	29.7 °C	0.079 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	24.926 %HR	22.2 %HR	-2.726 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.190 %HR	45.9 %HR	-3.290 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.926 %HR	73.3 %HR	3.374 %HR	± 5.0 %HR	± 1.9E+00 %HR