

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	1.805 °C	2.0 °C	0.195 °C	± 0.5 °C	± 4.0E-01 °C
4 °C	3.820 °C	3.9 °C	0.080 °C	± 0.5 °C	± 4.0E-01 °C
6 °C	5.876 °C	5.9 °C	0.024 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.992 °C	7.8 °C	-0.192 °C	± 0.5 °C	± 4.0E-01 °C