Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
0.0625 in	0.0625 in	0 in	$\pm 0.0001 in$	±	5.8E-04 in
0.1000 in	0.1000 in	0 in	$\pm 0.0001$ in	±	5.8E-04 in
0.1250 in	0.1250 in	0 in	$\pm 0.0001$ in	±	5.8E-04 in
0.2500 in	0.2500 in	0 in	$\pm 0.0001$ in	±	5.8E-04 in
0.2000 in	0.2000 in	0 in	$\pm 0.0001$ in	±	5.8E-04 in
0.3000 in	0.3000 in	0 in	$\pm 0.0001$ in	±	5.8E-04 in
0.4000 in	0.4000 in	0 in	$\pm 0.0001$ in	±	5.8E-04 in
0.5000 in	0.5000 in	0 in	$\pm 0.0001$ in	±	5.8E-04 in
0.6000 in	0.6000 in	0 in	$\pm 0.0001 in$	±	5.8E-04 in
0.8000 in	0.8000 in	0 in	$\pm 0.0001 in$	±	5.8E-04 in
1.0000 in	1.0000 in	0 in	$\pm 0.0001 in$	±	5.8E-04 in