

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
2.0000 in	2.0000 in	0 in	± 0.0002 in	±	5.8E-04 in
2.0625 in	2.0625 in	0 in	± 0.0002 in	±	5.8E-04 in
2.1000 in	2.1000 in	0 in	± 0.0002 in	±	5.8E-04 in
2.1250 in	2.1250 in	0 in	± 0.0002 in	±	5.8E-04 in
2.2000 in	2.2000 in	0 in	± 0.0002 in	±	5.8E-04 in
2.2500 in	2.2500 in	0 in	± 0.0002 in	±	5.8E-04 in
2.3000 in	2.3000 in	0 in	± 0.0002 in	±	5.8E-04 in
2.4000 in	2.4000 in	0 in	± 0.0002 in	±	5.8E-04 in
2.5000 in	2.5000 in	0 in	± 0.0002 in	±	5.8E-04 in
2.8000 in	2.8000 in	0 in	± 0.0002 in	±	5.8E-04 in
3.0000 in	3.0000 in	0 in	± 0.0002 in	±	5.8E-04 in