Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
5.0000 in	5.0000 in	0 in	± 0.0002 in	± 7.0E-02 in	
5.0625 in	5.0625 in	0 in	$\pm 0.0002$ in	± 5.7E-02 in	
5.1000 in	5.1000 in	0 in	$\pm 0.0002$ in	± 7.6E-02 in	
5.1250 in	5.1250 in	0 in	$\pm 0.0002$ in	± 9.4E-02 in	
5.2000 in	5.2000 in	0 in	$\pm 0.0002$ in	± 7.0E-02 in	
5.2500 in	5.2500 in	0 in	$\pm 0.0002$ in	± 9.9E-02 in	
5.3000 in	5.3000 in	0 in	$\pm 0.0002$ in	± 7.6E-02 in	
5.4000 in	5.4000 in	0 in	$\pm 0.0002$ in	± 9.4E-02 in	
5.5000 in	5.5000 in	0 in	$\pm 0.0002 in$	± 6.9E-02 in	
5.8000 in	5.8000 in	0 in	$\pm 0.0002 in$	± 1.1E-01 in	
6.0000 in	6.0000 in	0 in	$\pm 0.0002 in$	± 1.3E-01 in	