

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | E.M.P <i>(M.P.E)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|----------------------|------------------------------------|------------|
| 10.00 µl                 | 10 µl            | -0.0027 µl           | ± 0.12 µl            | ±                                  | 5.1E-02 µl |
| 49.98 µl                 | 50 µl            | 0.0193 µl            | ± 0.50 µl            | ±                                  | 5.1E-02 µl |
| 100.00 µl                | 100 µl           | 0.0045 µl            | ± 0.80 µl            | ±                                  | 5.1E-02 µl |