

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 15 °C                      | 14.800 °C                  | 14.9 °C            | 0.100 °C               | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 25 °C                      | 24.738 °C                  | 24.9 °C            | 0.162 °C               | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 50 °C                      | 51.120 °C                  | 51.4 °C            | 0.280 °C               | ± 1.0 °C                      | ± 1.1E+00 °C                         |
| 25 %HR                     | 23.280 %HR                 | 24.0 %HR           | 0.720 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.304 %HR                 | 52.0 %HR           | 1.696 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 75 %HR                     | 73.638 %HR                 | 76.0 %HR           | 2.362 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |