

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.031 °C	15.6 °C	0.569 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	24.988 °C	25.4 °C	0.412 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	49.897 °C	50.2 °C	0.303 °C	± 1.0 °C	± 4.0E-01 °C
25 %HR	24.323 %HR	20 %HR	-4.323 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.499 %HR	46 %HR	-4.499 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.335 %HR	71 %HR	-4.335 %HR	± 5.0 %HR	± 1.9E+00 %HR