

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	14.800 °C	15.0 °C	0.200 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	24.738 °C	24.8 °C	0.062 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	51.120 °C	51.5 °C	0.380 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	23.280 %HR	24.0 %HR	0.720 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.304 %HR	53.0 %HR	2.696 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	73.638 %HR	76.0 %HR	2.362 %HR	± 5.0 %HR	± 1.9E+00 %HR