

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.071 °C	0.1 °C	0.171 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	19.924 °C	20.2 °C	0.276 °C	± 1.0 °C	± 4.0E-01 °C
40 °C	39.899 °C	40.8 °C	0.901 °C	± 1.0 °C	± 4.0E-01 °C
25 %HR	24.926 %HR	30 %HR	5.074 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.190 %HR	51 %HR	1.810 %HR	± 5.0 %HR	± 1.9E+00 %HR
90 %HR	89.425 %HR	86 %HR	-3.425 %HR	± 5.0 %HR	± 1.9E+00 %HR