

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 0 °C                       | -0.071 °C                  | -0.4 °C            | -0.329 °C              | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 20 °C                      | 19.924 °C                  | 19.6 °C            | -0.324 °C              | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 40 °C                      | 39.899 °C                  | 39.3 °C            | -0.599 °C              | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 25 %HR                     | 24.926 %HR                 | 29 %HR             | 4.074 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 49.190 %HR                 | 46 %HR             | -3.190 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 90 %HR                     | 89.425 %HR                 | 89 %HR             | -0.425 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |