

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.071 °C	-0.2 °C	-0.129 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	19.924 °C	19.7 °C	-0.224 °C	± 1.0 °C	± 4.0E-01 °C
40 °C	39.899 °C	39.0 °C	-0.899 °C	± 1.0 °C	± 4.0E-01 °C
25 %HR	24.926 %HR	29 %HR	4.074 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.190 %HR	46 %HR	-3.190 %HR	± 5.0 %HR	± 1.9E+00 %HR
90 %HR	89.425 %HR	90 %HR	0.575 %HR	± 5.0 %HR	± 1.9E+00 %HR