

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.071 °C	0.6 °C	0.671 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	19.924 °C	20.5 °C	0.576 °C	± 1.0 °C	± 4.0E-01 °C
40 °C	39.899 °C	39.3 °C	-0.599 °C	± 1.0 °C	± 4.0E-01 °C
25 %HR	24.926 %HR	29 %HR	4.074 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.190 %HR	53 %HR	3.810 %HR	± 5.0 %HR	± 1.9E+00 %HR
90 %HR	89.425 %HR	89 %HR	-0.425 %HR	± 5.0 %HR	± 1.9E+00 %HR