

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 20 °C             | 19.924 °C         | 20.15 °C  | 0.226 °C      | ± 0.21 °C            | ± 4.0E-01 °C                |
| 25 °C             | 25.610 °C         | 25.99 °C  | 0.380 °C      | ± 0.21 °C            | ± 4.0E-01 °C                |
| 25 %HR            | 24.926 %HR        | 22.9 %HR  | -2.026 %HR    | ± 2.0 %HR            | ± 1.9E+00 %HR               |
| 50 %HR            | 49.190 %HR        | 47.7 %HR  | -1.490 %HR    | ± 2.0 %HR            | ± 1.9E+00 %HR               |
| 75 %HR            | 75.039 %HR        | 74.7 %HR  | -0.339 %HR    | ± 2.0 %HR            | ± 1.9E+00 %HR               |