

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	19.924 °C	20.07 °C	0.146 °C	± 0.21 °C	± 4.0E-01 °C
25 °C	25.610 °C	25.79 °C	0.180 °C	± 0.21 °C	± 4.0E-01 °C
25 %HR	24.926 %HR	23.1 %HR	-1.826 %HR	± 2.0 %HR	± 1.9E+00 %HR
50 %HR	49.190 %HR	48.0 %HR	-1.190 %HR	± 2.0 %HR	± 1.9E+00 %HR
75 %HR	75.039 %HR	73.2 %HR	-1.839 %HR	± 2.0 %HR	± 1.9E+00 %HR