

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.972 °C	29.91 °C	-0.062 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	33.423 %HR	35.25 %HR	1.827 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.302 %HR	50.93 %HR	1.628 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.560 %HR	72.48 %HR	1.920 %HR	± 5.0 %HR	± 1.9E+00 %HR