

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.052 °C	30.0 °C	-0.052 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.654 %HR	29.4 %HR	-2.254 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.499 %HR	48.0 %HR	-2.499 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.223 %HR	72.5 %HR	2.277 %HR	± 3.0 %HR	± 1.9E+00 %HR