

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
60 RPM	60.0 RPM	0.0 RPM	± 0.13 RPM	±	5.8E-02 RPM
600 RPM	600.0 RPM	0.0 RPM	± 0.40 RPM	±	5.8E-02 RPM
6000 RPM	5999 RPM	-1 RPM	± 4.0 RPM	±	5.8E-01 RPM
18000 RPM	17996 RPM	-4 RPM	± 10 RPM	±	8.0E-01 RPM