

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (M.P.E)	Incertidumbre (Uncertainty)	
2.0027 1	2 1	-0.0027 1	± 0.10 1	±	3.7E-01 µl
3.0036 1	3 1	-0.0036 1	± 0.10 1	±	3.9E-01 µl
4.0074 1	4 1	-0.0074 1	± 0.10 1	±	4.2E-01 µl
5.0099 1	5 1	-0.0099 1	± 0.10 1	±	4.5E-01 µl
10.0188 1	10 1	-0.0188 1	± 0.10 1	±	6.7E-01 µl
11.9250 1	12 1	0.0750 1	± 0.10 1	±	8.0E-01 µl
15.0258 1	15 1	-0.0258 1	± 0.10 1	±	9.5E-01 µl
16.0350 1	16 1	-0.0350 1	± 0.10 1	±	1.0E+00 µl