

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
100.0272 ml	100 ml	-0.0272 ml	± 1.0 ml	±	3.5E-01 µl
399.9194 ml	400 ml	0.0806 ml	± 3.2 ml	±	3.5E-01 µl
500.1103 ml	500 ml	-0.1103 ml	± 5.0 ml	±	3.5E-01 µl
800.1624 ml	800 ml	-0.1624 ml	± 8.0 ml	±	3.5E-01 µl
1000.6315 ml	1000 ml	-0.6315 ml	± 10.0 ml	±	3.5E-01 µl