

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.363 °C	30.3 °C	-0.063 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.243 %HR	35.7 %HR	0.457 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	50.4 %HR	0.372 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.335 %HR	72.9 %HR	2.565 %HR	± 3.0 %HR	± 1.9E+00 %HR