

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
4 °C	3.940 °C	4.0 °C	0.060 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	14.680 °C	14.8 °C	0.120 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.792 °C	30.0 °C	0.208 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.480 %HR	33.8 %HR	2.320 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.251 %HR	50.6 %HR	1.349 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	64.047 %HR	63.6 %HR	-0.447 %HR	± 3.0 %HR	± 1.9E+00 %HR