

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.313 °C	30.1 °C	-0.213 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	30.3 %HR	-0.270 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.049 %HR	49.8 %HR	-0.249 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.652 %HR	69.6 %HR	-1.052 %HR	± 3.0 %HR	± 1.9E+00 %HR