

Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)	
50.00 mmHg	50 mmHg	0.00 mmHg	± 5 mmHg	±	6.6E-01 mmHg
100.00 mmHg	100 mmHg	0.00 mmHg	± 5 mmHg	±	6.5E-01 mmHg
149.00 mmHg	150 mmHg	1.00 mmHg	± 5 mmHg	±	6.6E-01 mmHg
199.00 mmHg	200 mmHg	1.00 mmHg	± 5 mmHg	±	6.8E-01 mmHg
248.00 mmHg	250 mmHg	2.00 mmHg	± 5 mmHg	±	7.5E-01 mmHg
298.00 mmHg	300 mmHg	2.00 mmHg	± 5 mmHg	±	7.5E-01 mmHg